

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)		Incertidumbre (<i>Uncertainty</i>)
30 °C	30.714 °C	30.7 °C	-0.014 °C	±	1 °C	± 4.0E-01 °C
35 %HR	36.511 %HR	38.6 %HR	2.089 %HR	±	3 %HR	± 1.9E+00 %HR
50 %HR	52.799 %HR	51.0 %HR	-1.799 %HR	±	3 %HR	± 1.9E+00 %HR
70 %HR	73.433 %HR	71.3 %HR	-2.133 %HR	±	3 %HR	± 1.9E+00 %HR