

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09920	0.800	± 1.2	± 4.3E-01 mm					
0 a 0.5	0.49902	0.980	± 1.6	± 4.3E-01 mm					
0 a 1	0.99884	1.160	± 2	± 4.3E-01 mm					
0 a 2	1.99910	0.900	± 2.8	± 4.8E-01 mm					
0 a 3	2.99898	1.020	± 3.6	± 4.8E-01 mm					
0 a 4	3.99925	0.750	± 4.4	± 4.8E-01 mm					
0 a 5	4.99912	0.880	± 5.2	± 4.8E-01 mm					
0 a 6	5.99876	1.240	± 6	± 3.6E-01 mm					
0 a 7	6.99930	0.700	± 6.8	± 3.6E-01 mm					
0 a 8	7.99933	0.670	± 7.6	± 3.6E-01 mm					