

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.336 °C	2.4 °C	0.064 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.252 °C	8.3 °C	0.048 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.141 %HR	71.0 %HR	0.859 %HR	± 3.0 %HR	± 1.9E+00 %HR