

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.336 °C	2.2 °C	-0.136 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.252 °C	8.1 °C	-0.152 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	31.0 °C	0.938 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.141 %HR	71.0 %HR	0.859 %HR	± 3.0 %HR	± 1.9E+00 %HR