

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.336 °C	3.0 °C	0.664 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.252 °C	8.5 °C	0.248 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.6 °C	-0.462 °C	± 1.0 °C	± 4.0E-01 °C
70 %HR	70.141 %HR	68.9 %HR	-1.241 %HR	± 5.0 %HR	± 1.9E+00 %HR