

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.032 °C	29.2 °C	-0.832 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	36.0 %HR	5.430 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.713 %HR	54.0 %HR	3.287 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	73.0 %HR	2.552 %HR	± 5.0 %HR	± 1.9E+00 %HR