

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.3	0.29920	0.800	± 1.4	± 4.3E-01 mm					
0 a 0.6	0.59921	0.790	± 1.6	± 4.3E-01 mm					
0 a 0.9	0.89921	0.790	± 1.9	± 4.3E-01 mm					
0 a 1.2	1.19923	0.770	± 2.1	± 4.8E-01 mm					
0 a 1.5	1.49923	0.770	± 2.4	± 4.8E-01 mm					
0 a 1.8	1.79924	0.760	± 2.6	± 4.8E-01 mm					
0 a 2.1	2.09925	0.750	± 2.8	± 4.8E-01 mm					
0 a 2.4	2.39927	0.730	± 3.1	± 4.8E-01 mm					
0 a 2.7	2.69927	0.730	± 3.3	± 4.8E-01 mm					
0 a 3	0.73000	0.730	± 3.6	± 4.8E-01 mm					