

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.985 °C	2.1 °C	0.115 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.991 °C	3.5 °C	-0.491 °C	± 0.5 °C	± 4.0E-01 °C
80 %HR	80.673 %HR	82.3 %HR	1.627 %HR	± 3.0 %HR	± 1.9E+00 %HR
95 %HR	90.591 %HR	93.3 %HR	2.709 %HR	± 3.0 %HR	± 1.9E+00 %HR

