

| Patrón <i>(Standard)</i> | UBP <i>(UIT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 12.1 RPM                 | 12 RPM           | -0.1 RPM             | ± 1 RPM ±                   | 1.0E-01 RPM                        |
| 29.8 RPM                 | 30 RPM           | 0.2 RPM              | ± 1 RPM ±                   | 1.0E-01 RPM                        |
| 59.9 RPM                 | 60 RPM           | 0.1 RPM              | ± 1 RPM ±                   | 1.0E-01 RPM                        |