

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
12.1 RPM	12 RPM	-0.1 RPM	± 1 RPM	± 1.0E-01	RPM
29.8 RPM	30 RPM	0.2 RPM	± 1 RPM	± 1.0E-01	RPM
59.9 RPM	60 RPM	0.1 RPM	± 1 RPM	± 1.0E-01	RPM