

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.286 °C	2.0 °C	-0.286 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.202 °C	8.0 °C	-0.202 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.550 °C	25.0 °C	-0.550 °C	± 1.0 °C	± 4.0E-01 °C
60 %HR	60.529 %HR	55.0 %HR	-5.529 %HR	± 5.0 %HR	± 1.9E+00 %HR