

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.007 °C	9.7 °C	-0.307 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.068 °C	31.6 °C	-0.468 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.386 %HR	42.1 %HR	1.714 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.662 %HR	78.6 %HR	2.938 %HR	± 3.0 %HR	± 1.9E+00 %HR