

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.007 °C	9.6 °C	-0.407 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.068 °C	31.9 °C	-0.168 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.386 %HR	42.7 %HR	2.314 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.662 %HR	78.4 %HR	2.738 %HR	± 3.0 %HR	± 1.9E+00 %HR