

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 10 °C                      | 10.007 °C                  | 10.0 °C            | -0.007 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 32 °C                      | 32.068 °C                  | 31.7 °C            | -0.368 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 40 %HR                     | 40.386 %HR                 | 41.2 %HR           | 0.814 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 75.662 %HR                 | 77.3 %HR           | 1.638 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |