

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.770 °C	15.7 °C	0.930 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.714 °C	20.1 °C	0.386 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.798 °C	25.6 °C	0.802 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.072 °C	30.4 °C	0.328 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.212 %HR	34 %HR	3.788 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.629 %HR	55 %HR	5.371 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	82.155 %HR	89 %HR	6.845 %HR	± 5.0 %HR	± 1.9E+00 %HR