

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.770 °C	15.3 °C	0.530 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.714 °C	20.2 °C	0.486 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.798 °C	25.1 °C	0.302 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.072 °C	29.7 °C	-0.372 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.212 %HR	34 %HR	3.788 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.629 %HR	50 %HR	0.371 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	82.155 %HR	80 %HR	-2.155 %HR	± 5.0 %HR	± 1.9E+00 %HR