

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.163 °C	30.1 °C	-0.063 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	33.4 %HR	2.830 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	52.5 %HR	1.807 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.1 %HR	0.652 %HR	± 3.0 %HR	± 1.9E+00 %HR

