

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
35 °C	35.086 °C	35.5 °C	0.414 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.212 %HR	35.2 %HR	4.988 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.629 %HR	49.5 %HR	-0.129 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.071 %HR	70.3 %HR	-0.771 %HR	± 5.0 %HR	± 1.9E+00 %HR