

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.772 °C	30.0 °C	0.228 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.597 %HR	30.8 %HR	2.203 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.404 %HR	51.2 %HR	1.796 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.660 %HR	72.3 %HR	-2.360 %HR	± 3.0 %HR	± 1.9E+00 %HR