

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.072 °C	30.1 °C	0.028 °C	± 0.8 °C	± 4.0E-01 °C
30 %HR	30.212 %HR	33.8 %HR	3.588 %HR	± 4.0 %HR	± 1.9E+00 %HR
50 %HR	49.629 %HR	52.2 %HR	2.571 %HR	± 4.0 %HR	± 1.9E+00 %HR
70 %HR	71.071 %HR	72.4 %HR	1.329 %HR	± 4.0 %HR	± 1.9E+00 %HR