

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.222 °C	15.0 °C	-0.222 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.5 °C	0.151 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.448 %HR	68.1 %HR	-2.348 %HR	± 3.0 %HR	± 1.9E+00 %HR

