

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00 mm	1.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
1.25 mm	1.25 mm	0 µm	± 30 µm	±	5.8E+00 µm
1.50 mm	1.50 mm	0 µm	± 30 µm	±	5.8E+00 µm
2.00 mm	1.99 mm	-10 µm	± 30 µm	±	5.8E+00 µm
3.00 mm	3.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
5.00 mm	5.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
6.00 mm	6.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
8.00 mm	8.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
10.00 mm	10.01 mm	10 µm	± 30 µm	±	5.8E+00 µm