

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-10 °C	-10.369 °C	-9.8 °C	0.569 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.391 °C	0.2 °C	0.591 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.259 °C	25.9 °C	0.641 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.290 %HR	28.6 %HR	0.310 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.263 %HR	45.5 %HR	-4.763 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	74.353 %HR	69.3 %HR	-5.053 %HR	± 5.0 %HR	± 1.9E+00 %HR