

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-10 °C	-10.369 °C	-9.9 °C	0.469 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.391 °C	0.0 °C	0.391 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.269 °C	25.0 °C	-0.269 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.372 %HR	21.2 %HR	-7.172 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.333 %HR	41.4 %HR	-7.933 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	73.883 %HR	66.1 %HR	-7.783 %HR	± 5.0 %HR	± 1.9E+00 %HR