

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-10 °C	-10.469 °C	-10.0 °C	0.469 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.291 °C	0.1 °C	0.391 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.309 °C	25.0 °C	-0.309 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.484 %HR	25.3 %HR	-3.184 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.120 %HR	47.5 %HR	-2.620 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	73.627 %HR	69.1 %HR	-4.527 %HR	± 5.0 %HR	± 1.9E+00 %HR