

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.570 %HR	35.3 %HR	4.730 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	49.998 %HR	53.1 %HR	3.102 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	68.0 %HR	-2.652 %HR	± 5 %HR	± 1.9E+00 %HR
90 %HR	90.284 %HR	85.8 %HR	-4.484 %HR	± 5 %HR	± 1.9E+00 %HR

