

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 10 °C             | 10.107 °C         | 10.2 °C   | 0.093 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.149 °C         | 24.9 °C   | -0.249 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.062 °C         | 30.0 °C   | -0.062 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.391 °C         | 40.3 °C   | -0.091 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

