

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.107 °C	10.5 °C	0.393 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.149 °C	25.5 °C	0.351 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.8 °C	-0.262 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.391 °C	40.4 °C	0.009 °C	± 0.5 °C	± 4.0E-01 °C

