

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.570 %HR	34.2 %HR	3.630 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	49.998 %HR	52.9 %HR	2.902 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	67.0 %HR	-3.652 %HR	± 5 %HR	± 1.9E+00 %HR
90 %HR	90.284 %HR	87.3 %HR	-2.984 %HR	± 5 %HR	± 1.9E+00 %HR

