

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	21.960 °C	22.2 °C	0.240 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.918 °C	25.2 °C	0.282 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.737 °C	32.0 °C	0.263 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	46.347 %HR	49.0 %HR	2.653 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.073 %HR	57.6 %HR	0.527 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.728 %HR	65.3 %HR	-2.428 %HR	± 3.0 %HR	± 1.9E+00 %HR