

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	21.960 °C	21.8 °C	-0.160 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.918 °C	24.7 °C	-0.218 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.737 °C	31.6 °C	-0.137 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	46.347 %HR	49.3 %HR	2.953 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.073 %HR	56.4 %HR	-0.673 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.728 %HR	65.0 %HR	-2.728 %HR	± 3.0 %HR	± 1.9E+00 %HR