

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	21.960 °C	22.4 °C	0.440 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.918 °C	25.4 °C	0.482 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.737 °C	32.2 °C	0.463 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	46.347 %HR	49.2 %HR	2.853 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.073 %HR	56.0 %HR	-1.073 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.728 %HR	64.7 %HR	-3.028 %HR	± 3.0 %HR	± 1.9E+00 %HR