

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	21.960 °C	22.3 °C	0.340 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.918 °C	25.3 °C	0.382 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.737 °C	32.1 °C	0.363 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	46.347 %HR	49.5 %HR	3.153 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.073 %HR	56.3 %HR	-0.773 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.728 %HR	65.5 %HR	-2.228 %HR	± 3.0 %HR	± 1.9E+00 %HR