

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	21.960 °C	22.0 °C	0.040 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.918 °C	25.1 °C	0.182 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.737 °C	31.8 °C	0.063 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	46.347 %HR	49.3 %HR	2.953 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.073 %HR	55.5 %HR	-1.573 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.728 %HR	65.0 %HR	-2.728 %HR	± 3.0 %HR	± 1.9E+00 %HR