

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 130.3 μm          | 130.2 μm  | -0.1 μm       | ± 5 μm            | ±                           | 5.8E-01 μm |
| 523.0 μm          | 523.2 μm  | 0.2 μm        | ± 5 μm            | ±                           | 5.8E-01 μm |