

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
35 °C	35.126 °C	35.3 °C	0.174 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.110 %HR	30.9 %HR	0.790 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	51.6 %HR	0.907 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	72.3 %HR	2.159 %HR	± 3.0 %HR	± 1.9E+00 %HR