

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
25 °C	24.988 °C	24 °C	-0.988 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	33.954 %HR	33 %HR	-0.954 %HR	± 5.0 %HR	± 1.9E+00 %HR
90 %HR	88.996 %HR	86 %HR	-2.996 %HR	± 5.0 %HR	± 1.9E+00 %HR