

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.308 °C	10.28 °C	-0.028 °C	± 0.5 °C	± 1.8E-01 °C
20 °C	19.834 °C	19.41 °C	-0.424 °C	± 0.5 °C	± 1.8E-01 °C
30 °C	30.263 °C	30.04 °C	-0.223 °C	± 0.5 °C	± 1.6E-01 °C
40 °C	40.291 °C	39.97 °C	-0.321 °C	± 0.5 °C	± 1.6E-01 °C
50 %HR	50.263 %HR	52.84 %HR	2.577 %HR	± 3.0 %HR	± 1.2E+00 %HR
60 %HR	60.529 %HR	61.96 %HR	1.431 %HR	± 3.0 %HR	± 1.2E+00 %HR
70 %HR	70.448 %HR	75.53 %HR	5.082 %HR	± 3.0 %HR	± 1.2E+00 %HR
80 %HR	80.161 %HR	86.42 %HR	6.259 %HR	± 3.0 %HR	± 1.1E+00 %HR
90 %HR	90.703 %HR	100.00 %HR	9.297 %HR	± 3.000 %HR	± 1.1E+00 %HR
100 %HR	99.988 %HR	100.00 %HR	0.012 %HR	± 3.000 %HR	± 1.1E+00 %HR