

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.308 °C	10.28 °C	-0.028 °C	± 0.5 °C	± 2.4E-01 °C
20 °C	19.834 °C	19.41 °C	-0.424 °C	± 0.5 °C	± 2.4E-01 °C
30 °C	30.263 °C	30.04 °C	-0.223 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.291 °C	39.97 °C	-0.321 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	50.263 %HR	52.84 %HR	2.577 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	60.529 %HR	61.96 %HR	1.431 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	75.53 %HR	5.082 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.161 %HR	86.42 %HR	6.259 %HR	± 3.0 %HR	± 1.3E+00 %HR
90 %HR	90.703 %HR	100.00 %HR	9.297 %HR	± 3.0 %HR	± 1.3E+00 %HR
# 100 %HR	99.988 %HR	100.00 %HR	0.012 %HR	± 3.0 %HR	± 1.3E+00 %HR