

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.308 °C	10.28 °C	-0.028 °C	± 0.5 °C	± 1.8E-01 °C
20 °C	19.934 °C	19.94 °C	0.006 °C	± 0.5 °C	± 1.8E-01 °C
30 °C	30.263 °C	29.95 °C	-0.313 °C	± 0.5 °C	± 1.6E-01 °C
40 °C	40.291 °C	40.13 °C	-0.161 °C	± 0.5 °C	± 1.6E-01 °C
50 %HR	50.263 %HR	53.60 %HR	3.337 %HR	± 3.0 %HR	± 1.2E+00 %HR
60 %HR	60.529 %HR	62.43 %HR	1.901 %HR	± 3.0 %HR	± 1.2E+00 %HR
70 %HR	70.448 %HR	73.89 %HR	3.442 %HR	± 3.0 %HR	± 1.2E+00 %HR
80 %HR	80.059 %HR	84.33 %HR	4.271 %HR	± 3.0 %HR	± 1.1E+00 %HR
90 %HR	90.703 %HR	100.00 %HR	9.297 %HR	± 3.0 %HR	± 1.1E+00 %HR
100 %HR	99.988 %HR	100.00 %HR	0.012 %HR	± 3.0 %HR	± 1.1E+00 %HR