

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
35 °C	35.016 °C	35.3 °C	0.284 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.842 %HR	25.9 %HR	-2.942 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.507 %HR	45.4 %HR	-4.107 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	73.658 %HR	68.2 %HR	-5.458 %HR	± 5.0 %HR	± 1.9E+00 %HR