

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
35 °C	35.016 °C	34.5 °C	-0.516 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.842 %HR	25.8 %HR	-3.042 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.507 %HR	47.5 %HR	-2.007 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	73.658 %HR	72.7 %HR	-0.958 %HR	± 5.0 %HR	± 1.9E+00 %HR