

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
35 °C	35.126 °C	35.1 °C	-0.026 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.110 %HR	31.4 %HR	1.290 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	50.5 %HR	-0.193 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	67.8 %HR	-2.341 %HR	± 3.0 %HR	± 1.9E+00 %HR