

Certificado No. (Certificate #):		55624		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-04-04		Marca (Brand):		kynup	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		5 mm	5.00 mm	0 µm	± 30 µm	± 7.7E+00 µm	
Exterior / (Exterior)		30 mm	30.00 mm	0 µm	± 30 µm	± 7.8E+00 µm	
Exterior / (Exterior)		110 mm	110.00 mm	0 µm	± 30 µm	± 9.1E+00 µm	
Exterior / (Exterior)		220 mm	220.00 mm	0 µm	± 40 µm	± 1.2E+01 µm	
Exterior / (Exterior)		280 mm	280.00 mm	0 µm	± 40 µm	± 1.5E+01 µm	
Interior / (Interior)		50 mm	50.00 mm	0 µm	± 30 µm	± 8.5E+00 µm	
Interior / (Interior)		150 mm	150.00 mm	0 µm	± 30 µm	± 8.3E+00 µm	
Profundidad / (Depth)		80 mm	80.02 mm	23 µm	± 30 µm	± 7.9E+00 µm	
Profundidad / (Depth)		200 mm	200.03 mm	33 µm	± 30 µm	± 1.1E+01 µm	