

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.772 °C	30.0 °C	0.228 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.597 %HR	30.4 %HR	1.803 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.404 %HR	50.6 %HR	1.196 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.660 %HR	72.3 %HR	-2.360 %HR	± 3.0 %HR	± 1.9E+00 %HR