

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)		Incertidumbre (<i>Uncertainty</i>)
30 °C	30.042 °C	30.3 °C	0.258 °C	±	1 °C	± 4.0E-01 °C
35 %HR	33.995 %HR	35.3 %HR	1.305 %HR	±	3 %HR	± 1.9E+00 %HR
50 %HR	49.323 %HR	51.2 %HR	1.877 %HR	±	3 %HR	± 1.9E+00 %HR
70 %HR	73.259 %HR	75.7 %HR	2.441 %HR	±	3 %HR	± 1.9E+00 %HR