

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.089 °C        | -25.3 °C  | -0.211 °C     | ± 0.7 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.722 °C         | -0.8 °C   | -0.078 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 5.495 °C          | 5.1 °C    | -0.395 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.308 °C         | 10.0 °C   | -0.308 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.049 °C         | 24.7 °C   | -0.349 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.163 °C         | 29.8 °C   | -0.363 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.491 °C         | 40.0 °C   | -0.491 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

