

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.089 °C	-25.4 °C	-0.311 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.121 °C	-0.5 °C	-0.379 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.1 °C	0.107 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.1 °C	0.093 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.3 °C	-0.049 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.491 °C	40.5 °C	0.009 °C	± 0.5 °C	± 4.0E-01 °C

