

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.089 °C	-24.6 °C	0.489 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.020 °C	0.4 °C	0.420 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.0 °C	0.007 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.308 °C	10.1 °C	-0.208 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.1 °C	-0.350 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.664 °C	30.3 °C	-0.364 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.190 °C	40.2 °C	0.010 °C	± 0.5 °C	± 4.0E-01 °C

