

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.089 °C        | -25.5 °C  | -0.411 °C     | ± 0.7 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.020 °C         | -0.3 °C   | -0.280 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 4.993 °C          | 4.6 °C    | -0.393 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.308 °C         | 10.4 °C   | 0.092 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.450 °C         | 25.1 °C   | -0.350 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.664 °C         | 30.2 °C   | -0.464 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.190 °C         | 39.7 °C   | -0.490 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

